



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

Laboratory Report

Merrimack Village Dist
2 Greens Pond Rd
Merrimack, NH 03054

Date Printed: 01/07/2026
Work Order #: 2512-04813
Client Job #:
Date Received: 12/30/2025
Sample collected in: New Hampshire

Attached please find results for the analysis of the samples received on the date referenced above.

Unless otherwise noted in the attached report, the analyses performed met the requirements of the analyzing laboratory's Quality Assurance Plan, Standard Operating Procedures and State Accreditation. This certificate shall not be reproduced, except in full, without the written approval of the analyzing laboratory. The results presented in this report relate to the samples listed on the following pages in the condition in which they were received. Accreditation for each analyte is identified by the * symbol following the analyte name. Location of our analyzing laboratory is identified by the code in the Analyst Column.

A & L Laboratory:
Identified by ME in Analyst Column
155 Center Street, Auburn, Maine 04210
www.allaboratory.com

Granite State Analytical Services LLC:
Identified by NH in Analyst Column
22 Manchester Road, Derry, NH 03038
www.granitestateanalytical.com

Nashoba Analytical:
Identified by MA in the Analyst Column
31A Willow Road, Ayer, MA 01432
www.nashobaanalytical.com

ANALYSIS RELATED NOTES:

- RL: "Reporting limit" means the lowest level of an analyte that can be accurately recovered from the matrix of interest.
- DF: "Dilution factor" means the ratio of the volume of the sample to the volume of the final (dilute) solution.
- MDL: "Minimum Detection Limit" means the minimum result which can be reliably discriminated from a blank with a predetermined confidence level.
- ND: Non-detect. Results reported as Non-Detect (ND) have been evaluated down to the concentration listed in the MDL column.
- A & L Laboratory / Granite State Analytical Services LLC / Nashoba Analytical. accreditation lists can be found on our websites listed above.
- Subcontracted samples will be identified by the Accreditation number of the subcontract laboratory in the analyst field for each analyte and the appropriate laboratory will be listed here. **None**
- Data Qualifiers (DQ) Flags provide additional information in regards to the receipt, analysis or quality control of a sample. These are indicated under the DQ Flags Column on your report and listed here if necessary: **Data Qualifier (DQ) Flags: J = Estimated concentration.**

SAMPLE STATE SPECIFIC NOTES:

Additional Narrative or Comments: **None**

We appreciate the opportunity to provide you with laboratory services. If you have any questions regarding the enclosed report, please contact the laboratory and we will be happy to assist you.



Donald A. D'Anjou, Ph. D.
Laboratory Director

A & L Laboratory: Accreditations: Maine ME00021, New Hampshire 2501, Maine Radon Registration ID # SPC20
Granite State Analytical Services, LLC: Accreditations: New Hampshire 1015; Maine NH00003;
Massachusetts M-NH0003; Rhode Island 101513; Vermont VT-101507
Nashoba Analytical: Accreditations: Massachusetts M-MA1118



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-001
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-008 (MVD-2 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 11:25AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroicosafafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF30UdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF30NS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorobutanesulfonic Acid (PFBS)*	3.90	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorobutanoic Acid (PFBA)*	3.18	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-001
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-008 (MVD-2 raw)

EPA ID#: 1531010

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

DATE & TIME COLLECTED: 12/30/2025 11:25AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoroheptanoic acid (PFHpA)*	4.30	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorohexanesulfonic Acid (PFHxS)*	1.44	ng/L	✓	J	2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorohexanoic acid (PFHxA)*	5.33	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorooctanesulfonic acid (PFOS)*	3.28	ng/L	✓		2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluorooctanoic acid (PFOA)*	22.7	ng/L	✕		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoropentanoic acid (PFPeA)*	3.88	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:13PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	72	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	80	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	65	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	78	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-001
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-008 (MVD-2 raw)

EPA ID#: 1531010

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

DATE & TIME COLLECTED: 12/30/2025 11:25AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	82	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[13C8]octanoic acid	79	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Perfluoro-n-[13C9]nonanoic acid	80	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	83	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	121	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	70	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Sodium perfluoro-[13C8]octanesulfonate	87	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	91	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:13PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-002
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-012 (MVD 9 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 11:28AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF3OUdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorobutanesulfonic Acid (PFBS)*	2.02	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorobutanoic Acid (PFBA)*	1.21	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-002
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▼
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-012 (MVD 9 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 11:28AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoroheptanoic acid (PFHpA)*	1.00	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorohexanesulfonic Acid (PFHxS)*	<2.00	ng/L	✓		2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorohexanoic acid (PFHxA)*	1.09	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorooctanesulfonic acid (PFOS)*	1.20	ng/L	✓	J	2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluorooctanoic acid (PFOA)*	5.72	ng/L	✓		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoropentanoic acid (PFPeA)*	0.919	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:27PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	69	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	68	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	80	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
 Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-002
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-012 (MVD 9 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 11:28AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	88	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	82	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	87	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[13C8]octanoic acid	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Perfluoro-n-[13C9]nonanoic acid	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	121	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	72	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Sodium perfluoro-[13C8]octanesulfonate	87	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	91	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:27PM



Donald A. D'Anjou

Donald A. D'Anjou, Ph. D.
 Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-003
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-512 (TP-2/9 treated)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 11:37AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF30UdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorobutanesulfonic Acid (PFBS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorobutanoic Acid (PFBA)*	3.72	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-003
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-512 (TP-2/9 treated)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 11:37AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoroheptanoic acid (PFHpA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorohexanesulfonic Acid (PFHxS)*	<2.00	ng/L	✓		2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorohexanoic acid (PFHxA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorooctanesulfonic acid (PFOS)*	<2.00	ng/L	✓		2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluorooctanoic acid (PFOA)*	<2.00	ng/L	✓		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoropentanoic acid (PFPeA)*	1.94	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:41PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	58	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	67	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	64	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	67	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	66	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
 Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-003
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-512 (TP-2/9 treated)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 11:37AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	72	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	63	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	73	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[13C8]octanoic acid	60	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Perfluoro-n-[13C9]nonanoic acid	61	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	91	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	75	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Sodium perfluoro-[13C8]octanesulfonate	90	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	93	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	93	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:41PM



Donald A. D'Anjou

Donald A. D'Anjou, Ph. D.
 Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-004
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-011 (MVD-4R raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:15PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF30UdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	0.613	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF30NS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorobutanesulfonic Acid (PFBS)*	7.04	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorobutanoic Acid (PFBA)*	6.52	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-004
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-011 (MVD-4R raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:15PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoroheptanoic acid (PFHpA)*	14.0	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorohexanesulfonic Acid (PFHxS)*	3.32	ng/L	✓		2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorohexanoic acid (PFHxA)*	19.6	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorononanoic acid (PFNA)*	0.711	ng/L	✓	J	2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorooctanesulfonic acid (PFOS)*	5.07	ng/L	✓		2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluorooctanoic acid (PFOA)*	86.8	ng/L	✕		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoropentanesulfonic acid (PFPeS)*	0.725	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoropentanoic acid (PFPeA)*	12.7	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 04:54PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	61	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	68	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	69	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	62	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	68	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-004
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-011 (MVD-4R raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:15PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	77	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	71	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	74	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[13C8]octanoic acid	67	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Perfluoro-n-[13C9]nonanoic acid	68	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	83	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	98	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	61	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Sodium perfluoro-[13C8]octanesulfonate	88	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	94	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 04:54PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
 Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-005
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-005 (MVD-5 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:17PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF30UdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorobutanesulfonic Acid (PFBS)*	6.66	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorobutanoic Acid (PFBA)*	6.30	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM



Donald A. D'Anjou

Donald A. D'Anjou, Ph. D.
 Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-005
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▼
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-005 (MVD-5 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:17PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoroheptanoic acid (PFHpA)*	11.2	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorohexanesulfonic Acid (PFHxS)*	3.25	ng/L	✓		2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorohexanoic acid (PFHxA)*	16.0	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorooctanesulfonic acid (PFOS)*	2.54	ng/L	✓		2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluorooctanoic acid (PFOA)*	63.5	ng/L	✕		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoropentanoic acid (PFPeA)*	10.8	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:08PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	73	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	77	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	78	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	69	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	79	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-005
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▼
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-005 (MVD-5 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:17PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	82	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	82	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[13C8]octanoic acid	78	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Perfluoro-n-[13C9]nonanoic acid	77	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	112	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	68	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Sodium perfluoro-[13C8]octanesulfonate	89	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	94	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:08PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-006
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-509 (TP-4/5 treated)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:18PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF30UdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF30NS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorobutanesulfonic Acid (PFBS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorobutanoic Acid (PFBA)*	4.15	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-006
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▼
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-509 (TP-4/5 treated)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 12:18PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoroheptanoic acid (PFHpA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorohexanesulfonic Acid (PFHxS)*	<2.00	ng/L	✓		2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorohexanoic acid (PFHxA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorooctanesulfonic acid (PFOS)*	<2.00	ng/L	✓		2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluorooctanoic acid (PFOA)*	<2.00	ng/L	✓		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoropentanoic acid (PFPeA)*	0.879	ng/L		J	2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:22PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	61	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	70	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	67	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	65	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	66	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-006
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-509 (TP-4/5 treated)

EPA ID#: 1531010


DATE & TIME COLLECTED: 12/30/2025 12:18PM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	68	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	64	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	74	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[13C8]octanoic acid	62	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Perfluoro-n-[13C9]nonanoic acid	64	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	71	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Sodium perfluoro-[13C8]octanesulfonate	87	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	88	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	90	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:22PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-007
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-007 (MVD-7 raw)

EPA ID#: 1531010

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

DATE & TIME COLLECTED: 12/30/2025 10:35AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF30UdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF30NS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorobutanesulfonic Acid (PFBS)*	2.90	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorobutanoic Acid (PFBA)*	2.80	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-007
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▼
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-007 (MVD-7 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 10:35AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoroheptanoic acid (PFHpA)*	3.33	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorohexanesulfonic Acid (PFHxS)*	1.41	ng/L	✓	J	2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorohexanoic acid (PFHxA)*	2.93	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorooctanesulfonic acid (PFOS)*	3.36	ng/L	✓		2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluorooctanoic acid (PFOA)*	28.4	ng/L	✕		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoropentanoic acid (PFPeA)*	2.55	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:35PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	70	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	66	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	79	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-007
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-007 (MVD-7 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 10:35AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	88	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	87	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	87	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[13C8]octanoic acid	83	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Perfluoro-n-[13C9]nonanoic acid	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	146	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	74	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Sodium perfluoro-[13C8]octanesulfonate	89	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	94	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	83	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:35PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-008
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-009 (MVD-8 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 10:40AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF30UdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Date Extracted	-					No Limit	EPA 533	SH-NH	01/02/2026 08:00AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorobutanesulfonic Acid (PFBS)*	2.91	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorobutanoic Acid (PFBA)*	2.22	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM



Donald A. D'Anjou

Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-008
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-009 (MVD-8 raw)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 10:40AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoroheptanoic acid (PFHpA)*	3.60	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorohexanesulfonic Acid (PFHxS)*	1.31	ng/L	✓	J	2.00	18 ng/L	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorohexanoic acid (PFHxA)*	3.24	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorooctanesulfonic acid (PFOS)*	1.88	ng/L	✓	J	2.00	15 ng/L	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluorooctanoic acid (PFOA)*	27.3	ng/L	✕		2.00	12 ng/L	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoropentanoic acid (PFPeA)*	2.86	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/02/2026 05:49PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	70	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	80	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	68	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	79	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-008
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-009 (MVD-8 raw)

EPA ID#: 1531010

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

DATE & TIME COLLECTED: 12/30/2025 10:40AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	92	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	85	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	84	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[13C8]octanoic acid	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Perfluoro-n-[13C9]nonanoic acid	78	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	88	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	151	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	80	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Sodium perfluoro-[13C8]octanesulfonate	89	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	92	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	81	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/02/2026 05:49PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-009
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	
Fails EPA Primary	
Fails EPA Secondary	
Fails State Guideline	
Attention	

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-510 (TP-7/8 treated)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 10:49AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
11-Chloroeicosafluoro-3-oxaundecane-1-sulfonic Acid (11CI-PF3OUdS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
1H,1H,2H,2H-Perfluorodecane sulfonic acid (8:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
1H,1H,2H,2H-Perfluorohexanesulfonic acid (4:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
1H,1H,2H,2H-Perfluorooctane sulfonic acid (6:2FTS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
4,8-Dioxa-3H-perfluorononanoic acid (ADONA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
9-Chlorohexadecafluoro-3-oxanonane-1-sulfonic acid (9CI-PF3ONS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Date Extracted	-					No Limit	EPA 533	TF-NH	01/06/2026 08:08AM
Hexafluoropropylene oxide dimer acid (HFPO-DA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Nonafluoro-3,6-dioxaheptanoic acid (NFDHA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoro(2-ethoxyethane)sulfonic acid (PFEESA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoro-3-methoxypropanoic acid (PFMPA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoro-4-methoxybutanoic acid (PFMBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorobutanesulfonic Acid (PFBS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorobutanoic Acid (PFBA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-009
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-510 (TP-7/8 treated)

EPA ID#: 1531010

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▼
Fails State Guideline	✕
Attention	⚠

DATE & TIME COLLECTED: 12/30/2025 10:49AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluorodecanoic Acid (PFDA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorododecanoic acid (PFDoA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoroheptanesulfonic acid (PFHpS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoroheptanoic acid (PFHpA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorohexanesulfonic Acid (PFHxS)*	<2.00	ng/L	✓		2.00	18 ng/L	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorohexanoic acid (PFHxA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorononanoic acid (PFNA)*	<2.00	ng/L	✓		2.00	11 ng/L	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorooctanesulfonic acid (PFOS)*	<2.00	ng/L	✓		2.00	15 ng/L	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluorooctanoic acid (PFOA)*	0.910	ng/L	✓	J	2.00	12 ng/L	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoropentanesulfonic acid (PFPeS)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoropentanoic acid (PFPeA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
Perfluoroundecanoic acid (PFUnA)*	<2.00	ng/L			2.00	No Limit	EPA 533	SH-NH	01/06/2026 03:33PM
2,3,3,3-Tetrafluoro-2-(1,1,2,2,3,3,3-heptafluoropropoxy-13C3)propanoic acid	79	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[1,2,3,4,5,6,7-13C7]undecanoic acid	94	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[1,2,3,4,5,6-13C6]decanoic acid	96	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[1,2,3,4,5-13C5]pentanoic acid	91	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[1,2,3,4,6-13C5]hexanoic acid	90	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM



Donald A. D'Anjou, Ph. D.
Laboratory Director



GRANITE STATE ANALYTICAL SERVICES, LLC.

22 Manchester Road, Unit 2, Derry, NH 03038
Phone (800) 699-9920 | (603) 432-3044 website www.granitestateanalytical.com

DRINKING WATER NON COMPLIANCE REPORT

DATE PRINTED: 01/07/2026
SAMPLE ID #: 2512-04813-009
SAMPLED BY: Merrimack Village Dist

CHEMICAL RESULTS FOR THE QUARTER 2025
LAB ID#: 1015

Legend	
Passes	✓
Fails EPA Primary	⊗
Fails EPA Secondary	▽
Fails State Guideline	✕
Attention	⚠

SAMPLE CATEGORY: Routine Sample
SYSTEM NAME: Merrimack Village Dist
SYSTEM TOWN: Merrimack
SAMPLE AGENT #: 603-424-9241
SAMPLE LOCATION: 1531010-510 (TP-7/8 treated)

EPA ID#: 1531010

DATE & TIME COLLECTED: 12/30/2025 10:49AM
DATE & TIME RECEIVED: 12/30/2025 03:59PM
WATER SYS TYPE: Community
RECEIPT TEMP: ON ICE 8.8° CELSIUS
CLIENT JOB #:

BAR CODE:

Test Description	Result	Test Units	Pass /Fail	DQ Flag	RL	Limit	Method	Analyst	Date & Time Analyzed
Perfluoro-n-[1,2,3,4-13C4]butanoic acid	91	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[1,2,3,4-13C4]heptanoic acid	89	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[1,2-13C2]dodecanoic acid	96	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[13C8]octanoic acid	89	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Perfluoro-n-[13C9]nonanoic acid	92	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]decane sulfonate	99	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]hexane sulfonate	98	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Sodium 1H,1H,2H,2H-perfluoro-1-[1,2-13C2]octane sulfonate	87	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Sodium perfluoro-[13C8]octanesulfonate	95	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Sodium perfluoro-1-[1,2,3-13C3]hexanesulfonate	97	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM
Sodium perfluoro-1-[2,3,4-13C3]butanesulfonate	98	%	✓			50-200%	EPA 533 - IDA	SH-NH	01/06/2026 03:33PM



Donald A. D'Anjou, Ph. D.
Laboratory Director

